

### **DCS issues**

1. FD APD Monitor found frozen. This issue is propping most of the time. Shifters are restarting GUI.
2. During process of cooling dcm-2-01-10 (warm APD 59) the CSS GUI was getting frozen.
3. 24 V power supply for dcm-02-10-09 is yellow in color on power supply summary window. It is showing correct sense voltage, current etc.