

VME Intensity Monitor - Task #13918

Configure second ring pickup for PXIE

09/23/2016 10:07 AM - John Diamond

Status:	Assigned	Start date:	09/23/2016
Priority:	Normal	Due date:	
Assignee:	John Diamond	% Done:	100%
Category:	Deploy & Config	Estimated time:	3.00 hours
Target version:		Spent time:	0.50 hour
Description			
Brian hooked up a second ring pickup device to the second channel on the ring pickup digitizer. Configure with the same parameters as the first, create ACNET devices and update D150.			

History

#1 - 10/07/2016 01:02 PM - John Diamond

- Status changed from New to Assigned

- % Done changed from 0 to 100

Added configuration for the second ring pickup at PXIE to pxintstartup and created the following ACNET devices:

- P:M02EPC
- P:M02EPI
- P:M02EPA
- P:M02EPR

Added devices to D150 > PXIE > Sub-page 3