

VME Intensity Monitor - Necessary Maintenance #16553

NMLTOR is not responding after power outage

05/16/2017 10:58 AM - John Diamond

Status:	Closed	Start date:	05/16/2017
Priority:	High	Due date:	
Assignee:	John Diamond	% Done:	0%
Category:		Estimated time:	1.00 hour
Target version:		Spent time:	0.50 hour
Description			

History

#1 - 05/16/2017 11:18 AM - John Diamond

- Status changed from Assigned to Closed

Found the ethernet cable plugged into GigE port (which is the inactive port on the new 5500s).