

## adinstbpm - Task #18055

Milestone # 17719 (New): Booster operational test of one crate alongside existing BPM system

### No beam threshold

10/27/2017 08:51 AM - John Diamond

<b>Status:</b>	Resolved	<b>Start date:</b>	10/27/2017
<b>Priority:</b>	Normal	<b>Due date:</b>	
<b>Assignee:</b>	John Diamond	<b>% Done:</b>	100%
<b>Category:</b>		<b>Estimated time:</b>	2.00 hours
<b>Target version:</b>		<b>Spent time:</b>	1.00 hour
<b>Description</b>			
All of the other BPM systems have an intensity threshold that inhibits the position calculation.			

### History

#### #1 - 10/29/2017 09:58 AM - John Diamond

- Status changed from Assigned to Resolved
- % Done changed from 0 to 100

Implemented a threshold parameter in the BPM configuration section of bpmd.conf (the default is 50 counts). If the beam intensity is under the threshold we return NO\_BEAM (999.99).