

## adinstbpm - Task #23738

Milestone # 23736 (Closed): IOTA ACNET Interface Change Requests Dec 2019

### Change behavior of BPM TBT devices when readout is in progress

12/09/2019 02:36 PM - John Diamond

<b>Status:</b>	Closed	<b>Start date:</b>	12/09/2019
<b>Priority:</b>	Normal	<b>Due date:</b>	
<b>Assignee:</b>	Robert Santucci	<b>% Done:</b>	100%
<b>Category:</b>		<b>Estimated time:</b>	4.00 hours
<b>Target version:</b>		<b>Spent time:</b>	0.00 hour
<b>Description</b>			
<p>Currently we return a timeout error if a readout is in progress when ACNET attempts to read a TBT reading. If a readout is in progress then we should provide the next available reading.</p> <p>(Note that this could result in a scan of BPM devices returning data from different machine pulses if the ACNET reading is done asynchronously. Users should be advised to collect data on an event if possible)</p>			

#### History

#1 - 12/09/2019 02:51 PM - John Diamond

- Assignee set to Robert Santucci

#2 - 03/13/2020 08:28 AM - Robert Santucci

data will be reported to acnet from most recent or previous reading, users can correlate the index value to insure data is from the same reading

#3 - 03/13/2020 08:28 AM - Robert Santucci

- % Done changed from 0 to 100

- Status changed from New to Closed