

VME BPM - Feature #9006

Provide ACNET devices for signal magnitude on each BPM plate

06/02/2015 10:22 AM - John Diamond

Status:	Closed	Start date:	06/02/2015
Priority:	Normal	Due date:	
Assignee:	John Diamond	% Done:	100%
Category:		Estimated time:	5.00 hours
Target version:		Spent time:	5.00 hours
Description			

History

#1 - 06/02/2015 02:40 PM - John Diamond

- Status changed from New to Closed
- % Done changed from 0 to 100
- Estimated time changed from 8.00 h to 5.00 h

Implemented a ClassACNet accessor for magnitude signals. Created ACNET devices (S;pMxxx) for TG9 bpm node. Updated Wiki documentation with device descriptions. Created "mag" tab on S49 with new devices.